



# S.V. PRECISION INSTRUMENTS

Shed No. 11 A, 11 B, Type III, Industrial Estate, Kukatpally,  
Hyderabad - 500 072. TELANGANA, INDIA.

Ph: 040-23079359, 23075387, 23079485, 40245280, Mob: 08142174716  
E-mail: info@svprecision.com, ali@svprecision.com, svprecision@gmail.com

## Calibration Certificate

Format No. : F-04

Page No. 1 of 2

Note : This Report refers only to the Instruments supplied and may not be reproduced except in full without permission from Managing Partner, S.V. Precision Instruments, Hyderabad.	<b>Name &amp; Address of the Customer :</b>		Certificate No.	: 18/WL/4291	
	M/S. AA CALIBRATIONS PVT. LIMITED, WORKS-13C/18, VASUNDHARA, HAAT, VASUNDHARA, GHAZIABAD.		Date of Calibration	: 5-6-2018	
			MRC No. / Date	: 18/1195 / 5-6-2018	
			Calibration Valid upto (As per Customer's request)	: 4-6-2019	
	<b>Instrument Details</b>				
	Name	: MINILEVEL NT	Range	: $\pm 2$ mm/M	
	Make	: WYLER	Least Count	: 0.001 mm/M	
	Sr. No. / ID No.	: L0233 / AACPL/M/ML-01	Visual Inspection	: O.K	
	Zero Error	: Nil	Repeatability	: 0.001 mm/M	
	Environmental Condition	Room Temperature : $20 \pm 1^\circ\text{C}$ Humidity : 46.9% RH	Calibration Method	: CI - 031 Based on IS Method : ---	
Uncertainty of Measurement (95% confidence level for a coverage factor k=2) $\pm 4 \mu\text{m}$					
<b>Standard Equipments used for calibration</b>					
S. No	Instrument Name	Certificate No.	Valid Upto	Calibrated by	
1.	ELECTRONIC LEVEL (I0481 / SV/BL/095)	17100674/D1.02/C-164	07-11-2019	NPL-New Delhi	
2.	----				
3.	----				
4.	----				
5.	----				
6.	----				
➤ All reference standards are traceable to national standards					
Remarks	Nil				

Calibrated by  
(Venkatratnam)

**TRUE COPY**

Authorized Signatory  
*Ravinder Kumar*  
Quality Manager