



NUTAN CALIBRATION LAB.

NABL Accredited Laboratory



CALIBRATION CERTIFICATE

Certificate No.	NCL/CC0228	Page No.	
ULR No.	CC20011800000228F	1 of 1	
Company Name & Address	M/s. Future Brake Components Khasra No.-51/15, Swarn Park, Mundka, New Delhi-110041.	Date of Calibration	26.12.2018
		Suggested Due Date	26.12.2019
		Certificate Issue Date	27.12.2018
		UUC Condition	Good
		Calibration Performed At	Lab.

Detail of UUC Equipment:		Environmental Conditions	
Instrument Name	Digital Vernier Caliper	Temperature (Degree C)	
Range	0-200 mm	(20 ± 2)	
Least Count	0.01 mm	Relative Humidity (%)	
Make/Model No.	Truesize/-----	(50 ± 10)	
Serial No./I.D. No.	66918091/-----	Reference Standard	S.O.P. No.
Location	-----	IS: 3651-1982	NCL/SOP/M03

Standard Equipment Used

S.No.	Instrument Name	Make/Serial No.	Calibrated By	Certificate No.	Due Date
1	Caliper Checker	KCP/1548	Tanson Instrument	TI/C/CC/001/18	21/03/2020


Calibration Results

S.No.	Standard Value	UUC Value (External)	UUC Value (Internal)	UUC Value (Depth)
	(In mm)	(In mm)	(In mm)	(In mm)
1	20	19.98	19.99	20.00
2	50	49.97	49.99	50.00
3	100	99.97	99.99	99.99
4	150	149.96	149.99	149.99
5	200	199.96	199.99	-----

Note:-

1. The calibration results reported in this calibration certificate are valid at the time of & under stated condition.
2. This certificate cannot be reproduced except in full without our prior permission in writing.
3. This certificate refers only to the particular items submitted for calibration.
4. UUC- Unit under calibration.
5. Uncertainty Measurement at approx 95% confidence level and coverage factor k = 2 is ± 14.0 µm


(Gaurav Goswami)
Calibration Engineer




(Y.K. Sharma)
Technical Manager